

10964-8000. *Instrument:* Mass Spectrometer, Model VG 5400. *Manufacturer:* Fisons Instruments, United Kingdom. *Intended Use:* See notice at 60 FR 9662, February 21, 1995.

*Comments:* None received. *Decision:* Approved. No instrument of equivalent scientific value to the foreign instrument, for such purposes as it is intended to be used, is being manufactured in the United States. *Reasons:* The foreign instrument provides: (1) sensitivity equal to or greater than  $2.0 \times 10^{-4}$  A/torr at 800  $\mu$ A trap current and an air sample of 0.4 cc will produce a  $^3\text{He}$  count rate of at least 1700 cps and (2) the precision of measurement of the  $^3\text{He}/^4\text{He}$  ratio will be  $\pm 2.0\%$  ( $1\sigma$ ) on ten measurements of the 0.4 cc STP air sample as defined above.

These capabilities are pertinent to the applicant's intended purposes and we know of no other instrument or apparatus of equivalent scientific value to the foreign instrument which is being manufactured in the United States.

Frank W. Creel,  
Director, Statutory Import Programs Staff.  
[FR Doc. 95-10528 Filed 4-27-95; 8:45 am]  
BILLING CODE 3510-DS-F

#### **Simpson College, et al.; Notice of Consolidated Decision on Applications for Duty-Free Entry of Scientific Instruments**

This is a decision consolidated pursuant to Section 6(c) of the Educational, Scientific, and Cultural Materials Importation Act of 1966 (Pub. L. 89-651, 80 Stat. 897; 15 CFR part 301). Related records can be viewed between 8:30 A.M. and 5:00 P.M. in Room 4211, U.S. Department of Commerce, 14th and Constitution Avenue, N.W., Washington, D.C.

*Comments:* None received. *Decision:* Approved. No instrument of equivalent scientific value to the foreign instruments described below, for such purposes as each is intended to be used, is being manufactured in the United States.

*Docket Number:* 94-033R. *Applicant:* Simpson College, Indianola, IA 50125. *Instrument:* Rapid Kinetics Accessory, Model SFA-12. *Manufacturer:* Hi-Tech Scientific, United Kingdom. *Intended Use:* See notice at 59 FR 16188, April 6, 1994. *Reasons:* The foreign instrument provides designed capability for use in classroom demonstration of principles of stopped flow kinetics for reactions with half-lives of .5 to 5 seconds.

*Docket Number:* 94-071R. *Applicant:* University of Arkansas for Medical

Sciences, Little Rock, AR 72205-7122. *Instrument:* Rapid Kinetics Accessory, Model SFA-20. *Manufacturer:* Hi Tech Scientific, United Kingdom. *Intended Use:* See notice at 59 FR 31208, June 17, 1994. *Reasons:* The foreign instrument provides customized compatibility with the observation cell and a Hewlett Packard 8452A diode array spectrophotometer.

*Docket Number:* 94-131. *Applicant:* University of Rhode Island, Narragansett, RI 02882-1197. *Instrument:* 5-sample Anticoincidence Multicounter System, Model GM-25-5. *Manufacturer:* Riso National Laboratory, Denmark. *Intended Use:* See notice at 59 FR 60607, November 25, 1994. *Reasons:* The foreign instrument provides: (1) anti-coincidence mode detection to reduce background cosmic radiation, (2) low-level beta counting of less than 0.2 CPM, and (3) a 5-sample simultaneous measurement gas-flow multicounter unit.

*Docket Number:* 94-133. *Applicant:* Woods Hole Oceanographic Institution, Woods Hole, MA 02543. *Instrument:* Multi-Sensor Core Logger. *Manufacturer:* GEOTEK, United Kingdom. *Intended Use:* See notice at 59 FR 63762, December 9, 1994. *Reasons:* The foreign instrument provides measurements of bulk density, magnetic susceptibility, P-wave velocity and gamma ray absorption from extracted cores of sediment.

*Docket Number:* 94-138. *Applicant:* Texas A&M Research Foundation, College Station, TX 77843. *Instrument:* 600M Fluorometer and Accessories, Model AquatrackA MKIII. *Manufacturer:* Chelsea Instrument, United Kingdom. *Intended Use:* See notice at 59 FR 63762, December 9, 1994. *Reasons:* The foreign instrument provides: (1) chlorophyll-a measurements over 5 orders of magnitude using logarithmic scaling, (2) low power consumption and (3) deployment in seawater.

The National Institutes of Health advises in its memoranda of February 16, 1995 that (1) the capabilities of each of the foreign instruments described above are pertinent to each applicant's intended purpose and (2) they know of no domestic instrument or apparatus of equivalent scientific value for the intended use of each instrument.

We know of no other instrument or apparatus being manufactured in the United States which is of equivalent

scientific value to any of the foreign instruments.

Frank W. Creel,  
Director, Statutory Import Programs Staff.  
[FR Doc. 95-10527 Filed 4-27-95; 8:45 am]  
BILLING CODE 3510-DS-F

#### **Applications for Duty-Free Entry of Scientific Instruments**

Pursuant to Section 6(c) of the Educational, Scientific and Cultural Materials Importation Act of 1966 (Pub. L. 89-651; 80 Stat. 897; 15 CFR part 301), we invite comments on the question of whether instruments of equivalent scientific value, for the purposes for which the instruments shown below are intended to be used, are being manufactured in the United States.

Comments must comply with 15 CFR 301.5(a)(3) and (4) of the regulations and be filed within 20 days with the Statutory Import Programs Staff, U.S. Department of Commerce, Washington, D.C. 20230. Applications may be examined between 8:30 A.M. and 5:00 P.M. in Room 4211, U.S. Department of Commerce, 14th Street and Constitution Avenue, N.W., Washington, D.C.

*Docket Number:* 95-021. *Applicant:* Tulane University, 6823 St. Charles Avenue, New Orleans, LA 70118. *Instrument:* Ultra Sensitive Steady State Fluorimeter with Accessories, Model FS900CD. *Manufacturer:* Edinburgh Instruments Ltd., United Kingdom. *Intended Use:* The instrument will be used to monitor the behavior of organic and inorganic molecules by their emissive properties. Further, the instrument will be used to develop and study red sensitive dyes, a class of molecules which have attracted attention. In addition, the instrument will be used for training graduate and undergraduate students towards their degree. *Application Accepted by Commissioner of Customs:* March 28, 1995.

*Docket Number:* 95-022. *Applicant:* University of Illinois at Urbana-Champaign, Purchasing Division, 506 South Wright Street, Urbana, IL 61801. *Instrument:* Electron Microscope, Model CM200. *Manufacturer:* Philips, The Netherlands. *Intended Use:* The instrument will be used by pre and post-doctoral students as a research tool to examine the microscopic structure of the following specimens or compounds of interest: the brain, copolymers, inorganic compounds adrenocortical cells, the honey bee, FLAT-MAP proteins and sub-micron structures.