

The Department shall determine, and the Customs Service shall assess, antidumping duties on all appropriate entries. The Department will issue appraisal instructions directly to the U.S. Customs Service.

Furthermore, the following deposit requirements will be effective upon publication of the final results of this administrative review for all shipments of TRBs from Romania entered, or withdrawn from warehouse, for consumption on or after the publication date, as provided for by section 751(a)(2)(c) of the Act: (1) The cash deposit rate for TIE will be the rate we determine in the final results of review; (2) for the other companies named above which had no shipments during the POR and which were not found to have separate rates, Tehnoforestimportexport, Alexandria, Brasov, Barlad, Ploiesti, Slatina, and Suceava, and for all other Romanian exporters, the cash deposit rate will be 7.67%, the Romania-wide rate established in the most recent segment of the proceeding; and (3) for non-Romanian exporters of subject merchandise from Romania, the cash deposit rate will be the rate applicable to the Romanian supplier of that exporter. These deposit requirements, when imposed, shall remain in effect until publication of the final results of the next administrative review.

This notice serves as a preliminary reminder to importers of their responsibility under 19 CFR 353.26 of the Department's regulations to file a certificate regarding the reimbursement of antidumping duties prior to liquidation of the relevant entries during this review period. Failure to comply with this requirement could result in the Secretary's presumption that reimbursement of antidumping duties occurred and the subsequent assessment of double antidumping duties. This administrative review and notice are in accordance with section 751(a)(1) of the Act (19 U.S.C. 1675(a)(1)) and 19 CFR 353.22.

Dated: March 3, 1997.

Robert S. LaRussa,

*Acting Assistant Secretary for Import Administration.*

[FR Doc. 97-5895 Filed 3-10-97; 8:45 am]

BILLING CODE 3510-DS-P

### **Centers for Disease Control and Prevention; Decision on Application for Duty-Free Entry of Scientific Instrument**

This decision is made pursuant to Section 6(c) of the Educational, Scientific, and Cultural Materials

Importation Act of 1966 (Pub. L. 89-651, 80 Stat. 897; 15 CFR part 301). Related records can be viewed between 8:30 a.m. and 5 p.m. in Room 4211, U.S. Department of Commerce, 14th and Constitution Avenue, NW., Washington, D.C.

Docket Number: 96-116. Applicant: Centers for Disease Control and Prevention, Mailstop F17, 4770 Buford Hwy, N.E., Atlanta, GA 30341-3724. Instrument: Mass Spectrometer, Model VG AutoSpec. Manufacturer: Micromass, Ltd., United Kingdom. Intended Use: See notice at 61 FR 66017, December 18, 1996.

Comments: None received. Decision: Approved. No instrument of equivalent scientific value to the foreign instrument, for such purposes as it is intended to be used, is being manufactured in the United States. Reasons: The foreign instrument provides: (1) High sensitivity and resolving power, continuously variable to 80 000 (10% valley definition) in El mode and (2) extended mass range to 2000 at 8keV accelerating potential. The National Institutes of Health advises in its memorandum dated November 25, 1996 that (1) these capabilities are pertinent to the applicant's intended purpose and (2) it knows of no domestic instrument or apparatus of equivalent scientific value to the foreign instrument for the applicant's intended use.

We know of no other instrument or apparatus of equivalent scientific value to the foreign instrument which is being manufactured in the United States.

Frank W. Creel,

*Director, Statutory Import Programs Staff.*

[FR Doc. 97-6041 Filed 3-10-97; 8:45 am]

BILLING CODE 3510-DS-P

### **Federal Highway Administration; Decision on Application for Duty-Free Entry of Scientific Instrument**

This decision is made pursuant to Section 6(c) of the Educational, Scientific, and Cultural Materials Importation Act of 1966 (Pub. L. 89-651, 80 Stat. 897; 15 CFR part 301). Related records can be viewed between 8:30 a.m. and 5 p.m. in Room 4211, U.S. Department of Commerce, 14th and Constitution Avenue, NW, Washington, DC.

Docket Number: 96-124. Applicant: Federal Highway Administration, McLean, VA 22101-2296. Instrument: ACFM Crack Microgauge, Model U9. Manufacturer: Technical Software Consultants, Ltd., United Kingdom. Intended Use: See notice at 62 FR 979, January 7, 1997.

Comments: None received. Decision: Approved. No instrument of equivalent scientific value to the foreign instrument, for such purposes as it is intended to be used, is being manufactured in the United States. Reasons: The foreign instrument provides a 5 KHz drive coil and two directionally sensitive pickup coils employing alternating current field measurement to detect and size structural defects. A U.S. Department of Energy laboratory advised on February 11, 1997 that: (1) This capability is pertinent to the applicant's intended purpose and (2) it knows of no domestic instrument or apparatus of equivalent scientific value to the foreign instrument for the applicant's intended use.

We know of no other instrument or apparatus of equivalent scientific value to the foreign instrument which is being manufactured in the United States.

Frank W. Creel,

*Director, Statutory Import Programs Staff.*

[FR Doc. 97-5892 Filed 3-10-97; 8:45 am]

BILLING CODE 3510-DS-P

### **U.S. Geological Survey, Denver, et al.; Consolidated Decision on Applications for Duty-Free Entry of Scientific Instruments**

This is a decision consolidated pursuant to Section 6(c) of the Educational, Scientific, and Cultural Materials Importation Act of 1966 (Pub. L. 89-651, 80 Stat. 897; 15 CFR part 301). Related records can be viewed between 8:30 a.m. and 5 p.m. in Room 4211, U.S. Department of Commerce, 14th and Constitution Avenue, NW., Washington, DC.

Comments: None received. Decision: Approved. No instrument of equivalent scientific value to the foreign instruments described below, for such purposes as each is intended to be used, is being manufactured in the United States.

Docket Number: 96-127. Applicant: U. S. Geological Survey, Denver, CO 80225. Instrument: SIR Mass Spectrometer with Automated Sample Peripherals, Model Optima. Manufacturer: Micromass, United Kingdom. Intended Use: See notice at 62 FR 2133, January 15, 1997. Reasons: The foreign instrument provides: (1) An absolute sensitivity of 1 mass 44 ion per 1000 molecules of CO<sub>2</sub> and (2) an external precision of 0.02 for 10 bar µl of CO<sub>2</sub>.

Docket Number: 96-130. Applicant: State University of New York, Stony Brook, NY 11794. Instrument: Mass Spectrometer, Model Delta<sup>plus</sup>.